

Source: uhrad.com.

# Medical Imaging

“Seeing Through the Human Body”

Isabelle Bloch and Pietro Gori

isabelle.bloch@sorbonne-universite.fr, pietro.gori@telecom-  
paris.fr

September 2021

# History of medicine


- From humours...

Until 19th century western medicine was based on Galen's theory of humours, or principal fluids: black bile, yellow bile, phlegm (strange liquid related to tumors), and blood



- ...to auscultation

Stethoscope: Laennec in 1816



- ... and imaging



It all started in 1845 with the first  
X-ray image



Roentgen's wife



Von Kolliker

# History of Medical Imaging

## X Ray

- **1845**: Discovery of X-ray by **Wilhelm Conrad Röntgen**, german physicist.
- **1972**: First scanner set up by **Allan Mc Cornack** et **Godfrey N. Hounsfield**, Nobel prize in 1979 for this work.

## US

- **1915**: Ultrasound propagation principles for SONAR (SOund NAvigation Ranging).
- **1955**: First echography by **Inge Edler** (1911-2001), swedish cardiologist.

## MRI


- **1945**: Discovery of nuclear resonance under a magnetic field by **Edward Purcell** and **Felix Bloch**, both nobel prize winners in 1952.
- **1973**: First MRI image on an animal by the american chemist **Paul Lauterbur**. P. Lauterbur and P. Mansfield won the Nobel Prize for Physiology or Medicine in 2003 for the development of MRI.
- **1980**: Spectroscopy by magnetic resonance.

## Nuclear

- **1934**: Discovery of natural radioactivity by **Henri Becquerel**, **Pierre** et **Marie Curie**.  
Discovery of artificial radioactivity by **Irène et Frédéric Joliot-Curie**
- **1990**: Nuclear medicine development with scintigraphy and positron emission tomography.

# Terminology

- **Transmission:** Photons passing through the body
- **Absorption:** Partial or total absorption of energy in the patient
- **Scatter:** Radiation diverted in a new direction



Source: Utrecht University, NL

- **Axial** - plane normal to a vector from head to toe.
- **Coronal** - plane normal to a vector from front to back
- **Sagittal** - plane normal to a vector from left to right.
- **Oblique** – otherwise.

# Image types

- 2D:
  - A single slice.
  - A projection image.
- 3D:
  - A series of parallel slices with uniform spacing.
  - A series of slices with varying spacing and/or orientation.
  - A volumetric data set.

# Imaging Support



**Films  
(hard copy)**



**Digital Display  
(soft copy)**

# Imaging Modalities

- X-Ray:
  - Traditional
  - Computerized Tomography (CT)
- Ultrasound (US)
- Magnetic Resonance Imaging (MRI)
- Nuclear Imaging






**X-rays**



**CT**



**Ultrasound**



**MRI**

etc.

# A lot of differences

- Physical principles (important to correctly interpret the images!)
- Irradiation
- Resolution
- 2D / 3D
- Slice / projection
- Temporal sequences
- Differentiation of tissues
- Anatomy / function or metabolism
- Instrumentation
- ...


# Research in image processing and understanding for medical applications

- Reconstruction
- Improve image quality
- Organ detection, classification, segmentation, recognition
- Pathology detection and recognition
- Quantitative measurements
- Modeling
- 3D visualization
- Registration and information fusion
- Functional analysis, neurosciences
- Aid to diagnosis, therapeutical or surgical planning
- Follow-up
- ...


# Complex and multi-disciplinary domain

- Physics
- Mathematics
- Computer science
- Biology
- Medicine, anatomy
- Neuroscience, psychology
- Clinical studies and research
- Strong interaction and communication needed among actors
- And the patient ...

# Principle of X-ray imaging



# Electromagnetic spectrum



Source: Wikipedia


# Conventional X-Ray



# Conventional X-Ray

## Physics Principles

- Accelerating & decelerating electrons generate electromagnetic radiation
- Bombardment of a target material with a beam of fast electrons
- Electrons are emitted thermally from a heated cathode (C) and are accelerated toward the anode target (A) by the applied voltage potential  $V$  ( $\sim kV$ ).
- Electromagnetic radiation is produced by the **quick deceleration** of the electrons when hitting the target (**Bremsstrahlung**).
  - $\sim 99\%$  of energy converted into heat.
  - Few generated X-ray photons with wavelength set by the amount of energy loss  $E = h\nu = E_2 - E_1$






# Conventional X-Ray

## Image Formation


- X-ray beam goes through the patient, which decreases energy, depending on the absorption pattern inside the body.



# Conventional X-Ray


## Image Capture

- Image captured on phosphor screen, with conversion to visible light.
- Recording on film (negative image)
- Display on video monitor (positive image)




# Conventional X-Ray

## System Design



# Conventional X-Ray

## System Design



# Conventional X-Ray


- **Advantages:**
  - Fast.
  - Inexpensive.
- **Disadvantages:**
  - Ionizing radiation involved.
  - Projection images.
  - Poor blood vessel and soft-tissue contrasts.

# Conventional X-Ray

- **Applications:**
  - Lungs, breasts.
  - Gastro-intestinal track.
  - Blood vessels (contrast agent).
  - Bones, joints.

# Conventional X-Ray


Examples: Bones



# Conventional X-Ray

## Examples: Lungs

Patient with hemo-pneumothorax



Patient with pacemaker





# Conventional X-Ray

## Examples: Mammography

- Low kV can be used (little attenuation) => sufficient photoelectric absorption to differentiate between normal tissue and pathology
- Dose must be limited.
- Breast must be compressed
  - To avoid motion.
  - To minimize focus/film distance.
  - To have equal tissue thickness.





# Conventional X-Ray

## Examples: Fluoroscopy


- Real-time X-ray imaging.
- Direct visualization on screen.
- Used for guidance of interventions (mainly vascular and orthopedic).

Ex: Fluoroscopic guidance with umbilical catheter injecting contrast agent in the small bowel.





# Conventional X-Ray

## Examples: Fluoroscopy



Source: Utrecht University, NL

# Computerized Tomography (CT)



# CT

## Physics Principles

- Same as X-ray.

# CT

## Image Formation


- Same as X-ray.

# CT


## Image Capture

- Generation of a sliced view of body interior (“T”) (cf. tomos = slice in Greek).
- Computational intensive image reconstruction (“C”).

Measure = integral attenuation



$\mu(x,y) ?$




Patient's  
body

X-ray source

# CT

## Image Capture

- Translation and rotation of integral measure of absorption.
- Reconstruction: see next lecture






# CT

## Image Capture: Hounsfield Units

- Relates the linear attenuation coefficient of a local region  $\mu$  to the linear attenuation coefficient of water,  $\mu_w$  @ 70 keV ( $=E_{\text{eff}}$ ).
- Based on measurements with the original EMI scanner invented by Hounsfield.




- Assign calibrated values to gray scale of CT images -> allows for thresholding!


# CT

## System Design


- **1<sup>st</sup> generation:** Original resolution:  $80 \times 80$  pixels (ea.  $3 \times 3 \text{ mm}^2$ ), 13mm slice thickness, 5min scan time, 20min reconstruction.
- **2<sup>nd</sup> generation:** reduced number of view angles  $\Rightarrow$  scan time  $\sim 30$  s.
- **3<sup>rd</sup> generation:** scan time  $\sim$ seconds (reduced dose, motion artifacts). Reconstruction time  $\sim$ seconds.



Translation  
& Rotation



Translation  
& Rotation




only  
Rotation

# CT


## System Design

- **4<sup>th</sup> generation:** scan time, reconstruction time ~sec.



only Rotation of sources


- **Spiral:** Continuous linear motion of patient table during multiple scans. Increased coverage volume per rotation.



# CT

## Example: Image quality

1975



128x128 pixels, 1-4 hours acquisition,  
1-5 days computation


2000



512x512 pixels, 0.35 sec acquisition,  
<1sec computation


# CT

## Example: Lungs



# CT

## Example: Lungs



Ex: CT of right lung with active pulmonary tuberculosis: shows centrilobular lesions (nodules or branching linear structures 2-4 mm in diameter) in and around the small airways.

Source: [www.mypacs.net](http://www.mypacs.net)

# CT



## Examples: Abdomen






Diagnosis of appendicitis

# CT

## Example: Abdomen



Source: [www.mypacs.net](http://www.mypacs.net)






# CT

## Example: Blood vessels




# CT


## Example: Blood Vessels



# CT




# Ultrasound



# Ultrasound

## Physics Principle

- Traveling pressure waves through a medium.






# Ultrasound

## Image Formation


- Sound waves propagate through the body.
- At each interface, part is reflected, part is transmitted.
- Reflected waves are recorded back.



# Ultrasound

Ultrasound Pulse

Image Formation






# Ultrasound


## Image Capture

- **Recorded signals:**
  - Specular reflection, Diffuse scattering, Rayleigh scattering
- **Frame rate (FR)** = number of times per second a sweep of the ultrasound beam is performed.
  - The higher the frame rate, the better the temporal resolution.
  - Limited by:
    - Time delay from maximum depth.
    - Speed of sound.




# Ultrasound

## System Design




## Phased array transducer



Matching Layer

## Display



# Ultrasound

- **Advantages:**
  - Portable.
  - Inexpensive.
  - Good spatial resolution (<1mm).
  - High temporal resolution.
- **Disadvantages:**
  - Poor image quality.
  - High level of noise.

# Ultrasound

## Applications

- Obstetric
- Heart
- Gastro-intestinal track (liver, gall bladder, pancreas)
- Urinary track (bladder, kidneys)
- Genital organs (prostate, testicles, ovaries, uterus).
- Blood vessels (Doppler).

# Ultrasound

## Example: Fetus

- Establish number of fetus.
- Check position of placenta.
- Determine fetal age and development.
- Check for congenital defects.
- Check fetus position.



Fetus at 9 weeks



# Ultrasound

## Example: Obstetric

1961: First image of fetus





1990: First 3D visualization of fetus



# Ultrasound




## Example: Heart

- Valve diseases.
- Cardiac function (pumping efficiency).
- Blood clots in cardiac chamber.



# Ultrasound

## Example: Heart



- Real-Time 3D Ultrasound



# Ultrasound


## Examples: Abdomen

- Biopsy guidance.
- Diagnosis of abdominal pains, kidney stones, inflamed appendicitis.

kidney



liver




appendicitis

# Ultrasound

## Example: Blood vessels


- Artery blockage from blood clots.
- Arteriosclerosis (plaques).
- Stent position




# Ultrasound

## Example: Blood vessels

- Doppler




# Magnetic Resonance Imaging (MRI)




# MRI

## Physics Principle



No magnetic field



Magnetic field




- Nuclei with magnetic dipoles (H).
- Magnetic field  $B_0$  makes nuclei precess along its direction.
- The precession rate is the Larmor frequency.
- $f_L = \gamma B_0$  ( $43 \times 1.5 = 64$  MHz for Hydrogen into a 1.5T magnet).




# MRI

## Image Formation



- For every 1,000,000 nuclei in upper state ( $m_z = +1/2$ ) there are 1,000,001 nuclei in lower state ( $m_z = -1/2$ ).
- This difference is enough to result in macroscopic **net magnetization of material**.


- Nuclei with magnetic dipole moment  $\mu$ , and angular momentum  $\mathbf{S}$ .
- Net magnetization  $\mathbf{M}$  = density of  $\mu$ .
- Tip  $\mathbf{M}$  away from  $\mathbf{B}_0 \Rightarrow$  Precession at frequency  $\gamma B_0$ , producing a measurable **RF magnetic field**.



# MRI

## Image Capture


- As the magnetization precesses it creates its own RF magnetic field.
- This field is much smaller than the exciting RF field.
- It can be detected with a standard radio receiver (e.g. RF coils used to tip  $M$  away from  $B_0$ )
- The resulting signal from precession is called the **Free Induction Decay** or FID.



# MRI



## System Design

- **Main Magnet**
  - High, constant, Uniform Field,  $B_0$ .
- **Gradient Coils**
  - Produce pulsed, linear gradients in this field:  $G_x$ ,  $G_y$ , &  $G_z$
  - spatial localization
- **RF coils**
  - Transmitter:  $B_1$  Excites NMR signal to generate Free Induction Decay (FID).
  - Receiver: Senses FID.






Retour de l'aimantation à son état d'équilibre:




Enregistrement d'un courant dans la bobine de réception:



# MRI

## System Design



Magnet + shielding + cooling systems

gantry

# MRI

- **Advantages:**
  - High image quality.
  - Good spatial resolution (~1mm).
  - Almost all tissues.
  - Can provide information on both anatomy and function.
- **Disadvantages:**
  - Very expensive scanners.
  - Some patients cannot be imaged (pacemaker, claustrophobia).

# MRI


## Applications

- Brain (Multiple sclerosis, Alzheimer's disease, epilepsy, tumors)
- Spine
- Muscles, tendons.
- Blood vessels: angiography (strokes, aneurism)
- Heart
- Soft tissues



# MRI

## Example: Heart

- CINE MRI




Courtesy of Robert R. Edelman



# MRI


## Example: Heart



**MRI for myocardial perfusion**

# MRI

## Example: Blood vessels



## Magnetic Resonance Angiography

# MRI

## Example: Blood vessels




**Aorta**




# MRI


## Example: Brain



SAGITTAL



AXIAL






CORONAL

# MRI

## Example: Brain

- Diagnosis of Alzheimer's and Parkinson's disease.




# MRI

## Example: Blood vessels





## MR Angiography



# fMRI

## Functional neuroimaging modalities






neuro


capillar

↑ O<sub>2</sub> metabolism  
↑ blood flow  
↑ blood volume




dHb = deoxyhemoglobin  
**paramagnetic**  
HbO<sub>2</sub> = oxyhemoglobin  
**diamagnetic**


# Diffusion MRI



# Nuclear Imaging (SPECT, TEP)



# Electromagnetic spectrum



Source: Wikipedia




# Nuclear Imaging

## Physics Principle

Nuclear imaging is the use of **radioactive materials** for imaging structures and **function** inside the body.

Stability of Atomic Nuclei

- *ground state*: most stable.
- *metastable state*: nucleus is unstable, has long lifetime before transforming to another state.
- *excited state*: nucleus very unstable, exists in this state for small time period before transforming.




# Nuclear Imaging

Energy (in keV)                      Half-life (in hours)

Technétium $^{99m}\text{Tc}$	140	6
Iode $^{123}\text{I}$	159	13
Thallium $^{201}\text{Tl}$	75 et 135	73
Xénon $^{133}\text{Xe}$	81	127
Indium $^{111}\text{In}$	173 et 247	67

# Nuclear Imaging

## System Design



PM: photo multiplier

**Photo-electric effect**  
+ Compton (undesirable)

# Nuclear Imaging

## **SPECT (single photon emission computerized tomography)**

- Several projections
- Reconstruction (as for CT, see next lecture)
- Attenuation correction

# Nuclear Imaging






## PET (positron emission tomography)

- In unstable nucleus, (more protons than neutrons) proton converts to neutron under emission of **positron**.
- Positron travels limited distance in tissue.
- Positron can combine with electron of nearby atom and emits two 511 keV X-ray photons ( $\gamma$  rays) that travel in opposite directions.

	$^{15}\text{O}$	$^{13}\text{N}$	$^{11}\text{C}$	$^{18}\text{F}$
Half-life (in min)	2	10	20	110

# Nuclear Imaging

Positron Ring Detector



CRT monitor

(see next lecture)

# Nuclear Imaging

- **Advantages**

- Functional imaging: metabolism, immunology, function, flux (not an anatomical imaging technique!).

- **Disadvantages**

- Poor spatial resolution (a few mm).
- High level of noise.
- Expensive hardware.
- Need of cyclotron.

# Nuclear Imaging

## Applications

- Heart
- Lungs
- Kidneys
- Bladder
- Gastro-intestinal Track
- Endocrine system (hormones)
- Immune system (white blood cells)
- Tumors (ex: FDG)




# Nuclear Imaging

## New developments

- Detection using semi-conductors
- Combined machines:
  - PET/CT
  - PET/MRI


# Nuclear Imaging

## Whole Body



**Figure 1.** Whole-body gallium-67 scan showing lymphoma in the chest, the spleen, the abdomen, and the groin. The chest and spleen sites were known before the scan. The abdominal and groin uptakes were detected only by the scan.

Source: <http://www.nucmednet.com/galymph.htm>



**Clinical History:** A 71-year-old male with history of lung carcinoma, now with bilateral pulmonary nodules and mediastinal adenopathy. **Findings:** Bone Scan: Increased radiopharmaceutical uptake along cortical margins of the bilateral femurs and tibias. No evidence of bony metastases.


**Diagnosis:** Hypertrophic osteoarthropathy

<http://www.uhrad.com/spectarc.htm>

# Nuclear Imaging

## Example: Lung

SPECT scan for lung perfusion (pulmonary embolism)




Source: Brigham & women's Hospital

# Nuclear Imaging


## Example: Heart

- Perfusion Imaging




# Nuclear Imaging

## Example: Brain



# Nuclear Imaging


## Example: Brain



# Nuclear Imaging

## Example: Brain

PET with FDG tracer for diagnosis of brain tumor




Source: V. Cunningham, GlaxoSmithKline.

# Nuclear Imaging

## Example: Brain

PET for in-vivo testing of new drugs








# Optical imaging


Caliper

0




# **Research in Medical Imaging: a few examples**

## Registration of MRI and PET abdomen images




## Segmentation of cortical brain structures




## Segmentation of sub-cortical structures


Lateral ventricles  
Third ventricle  
Caudate nucleus  
Thalamus




## Segmentation of sub-cortical structures and tumors




# Diffusion Tensor MRI for Brain White Matter Fibers



# Functional Brain MRI




Distance entre doigt ~ 0.9 cm  
Distance I - V ~ 1.5 cm




*petit doigt*  
*majeur*  
*index*  
*pouce*

**Motor control of fingers**







## Functional MRI




**Motor control of right hand**

# Fetal Imaging




# Fetal Imaging




# Books

• **Foundation of Medical Imaging**, Z.H. Cho, J.P. Jones, M. Singh, John Wiley & Sons, Inc., New York 1993, ISBN 0-471-54573-2.



• **Principles of Medical Imaging**, K.K. Shung, M.B. Smith, B. Tsui, Academic Press, San Diego 1992, ISBN 0-12-640970-6.

• **The Essential Physics of Medical Imaging (2nd Edition)**, Jerrold T. Bushberg, J. Anthony Seibert, Edwin M. Leidholdt Jr., John M. Boone, Lippincott Williams & Wilkins; 2nd edition, ISBN 0683301187.



• **Image Reconstruction in Radiology**, J.A. Parker, CRS Press, Boca Raton, FL 1990, ISBN 0-8493-0150-5/90.

• **Principles of Magnetic Resonance Imaging**, Zhi-Pei Liang, P.C. Lauterbur IEEE Press, New York, NY 2000, ISBN 0-7803-4723-4.

